



Stanford Research Systems, Inc.
1290 C Reamwood Avenue, Sunnyvale, CA 94089, U.S.A.
Phone: 408-744-9040 FAX: 408-744-9049
Email: janied@thinkSRS.com .. Website: www.thinkSRS.com

10/8/2019

Volatility Statement for SR560, SR570

The SR560/570 contain various types of memories, as listed below:

Non-Volatile Memory:

16KB x 8 CMOS EPROM Not user modifiable;
Contains instrument's firmware and the factory
calibration bytes

Volatile Memory:

2KB x 8 SRAM user modifiable; used for last operation/user settings.
To clear:
Depress the RESET key while powering ON the unit (for SR560)
Depress the FILTER RESET key while powering ON the unit (for
SR570)

Or, remove the back-up lithium battery

Please feel free to contact SRS should you have any further questions.

Janie Du
Sales & Marketing