

Stanford Research Systems, Inc.

1290 C Reamwood Avenue, Sunnyvale, CA 94089 Phone: 408-744-9040 FAX: 408-744-9049

Email: janied@thinkSRS.com .. Website: www.thinkSRS.com

Volatility Statement for SR630

The SR630 contains a Static RAM which holds calibration bytes and instrument settings. It is backed up by a 3V lithium battery.

Location	Type	Size	Function	User	Data Input	To clear
				Modifiable?	Method	User Settings
U 102	Static	32K x 8	Instrument	Yes	Front Panel	Power ON with
	RAM	70 nS	Settings		RS-232	BKSP *
			and Cal		GPIB	Or, remove the
			Bytes			battery (BT100)

^{*} With power off, holding down BKSP button while powering on the instrument (clear instrument settings and bring back factory default cal bytes.)

Please feel free to contact SRS should you have any further questions.